



## WORLD MEDICO | Sensitive

- Black Sensitive Nitril
- Godkendt til kontakt med fødevarematerialer
- Godkendt til medicinsk brug
- Puddorfri



KATEGORI	EN-STANDARD
 CAT III	EN   

MATERIALER	FORPAKNING
100% Nitril	100 stk. pr. æske

Vare nr.	EAN	BRAND	MODEL	STØRRELSER
10013	6975390730865	WORLD	SENSITIVE	Small
10014	6975390730872	WORLD	SENSITIVE	Medium
10015	6975390730889	WORLD	SENSITIVE	Large
10016	6975390730896	WORLD	SENSITIVE	XL

Disse første classes 4 grams handsker, giver en unik beskyttelse mod bla. kemikalier. Handskerne sikrer høj komfort og pasform. Undersøgelseshandskerne meget slidstærke men er samtid ekstremt bløde og elastiske . De tilpasser sig håndens form under brug, og giver en god gennemfølgelse. Handskerne er åndbare men uigennemtrængelige. Handskerne sikrer et solidt greb og kan bruges til en lang række opgaver såsom rengøringsopgaver, plejeopgaver, håndtering af fødevarer, laboratorieopgaver, medicinhåndtering, cytotatika og hormoncremer. Indeholder ingen pudder eller latexproteiner og kan derfor sikkert anvendes ved latexallergi

God og kraftig engangshandske med rullekant.

Tykkelse:

Håndflade: 0,07 mm (+/- 0,02)

Fingre: 0,09 mm (+/- 0,02)

Manchet: 0,05 mm (+/- 0,02)

### Leverandør

Firma: World Medico ApS  
 Adresse: Østerfælled Torv 3  
 Post nr. / By: 2100 / København Ø  
 Land: Danmark  
 Telefon / E-mail: +45 53692712 / info@worldmedico.dk

<b>CE2777</b> <b>CAT III</b>	These products are compliant as Medical Device Class I, Category III Personal Protective Equipment (PPE) by the European PPE Regulation 2016/425 and have been shown to comply with this Regulation through the Harmonised European Standard(s): EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018, EN ISO 374-5:2016		EN ISO 374-1:2016/Type B														
	EN ISO 374-1:2016+A1:2018 Permeation levels are based on breakthrough times as follows:	<table border="1"> <thead> <tr> <th>Measured Breakthrough Time (min)</th> <th>Permeation Performance Level</th> </tr> </thead> <tbody> <tr> <td>&gt; 30min</td> <td>1</td> </tr> <tr> <td>&gt; 30min</td> <td>2</td> </tr> <tr> <td>&gt; 60min</td> <td>3</td> </tr> <tr> <td>&gt; 120min</td> <td>4</td> </tr> <tr> <td>&gt; 240min</td> <td>5</td> </tr> <tr> <td>&gt; 480min</td> <td>6</td> </tr> </tbody> </table>	Measured Breakthrough Time (min)	Permeation Performance Level	> 30min	1	> 30min	2	> 60min	3	> 120min	4	> 240min	5	> 480min	6	ISO 374-5:2016, class 2. The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen.  Before usage, inspect the gloves for any defector specifications. Components used in glove manufacturing may cause allergic reactions. If reactions occur then seek medical advice immediately. Only use with dry and clean hands.
Measured Breakthrough Time (min)	Permeation Performance Level																
> 30min	1																
> 30min	2																
> 60min	3																
> 120min	4																
> 240min	5																
> 480min	6																
EU Type Examination and On-going Conformity Module CZ. Carried out by Noti of Body 2772. SATRA Technology Europe Ltd. Bracetown Business Park, Clonsilla, Dublin 15, D15 YK2E Ireland	<table border="1"> <thead> <tr> <th>EN ISO 374-1:2016+A1:2018 /Type B</th> <th>Level</th> <th>EN ISO 374-4:2019 Degradation (%)</th> </tr> </thead> <tbody> <tr> <td>40% Sodium Hydroxide (O)</td> <td>6</td> <td>-5.9</td> </tr> <tr> <td>30% Hydrogen Peroxide (P)</td> <td>2</td> <td>36.4</td> </tr> <tr> <td>37% Formaldehyde (F)</td> <td>6</td> <td>-2.1</td> </tr> </tbody> </table>		EN ISO 374-1:2016+A1:2018 /Type B	Level	EN ISO 374-4:2019 Degradation (%)	40% Sodium Hydroxide (O)	6	-5.9	30% Hydrogen Peroxide (P)	2	36.4	37% Formaldehyde (F)	6	-2.1	EN ISO 374-4:2019 Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.		
EN ISO 374-1:2016+A1:2018 /Type B	Level	EN ISO 374-4:2019 Degradation (%)															
40% Sodium Hydroxide (O)	6	-5.9															
30% Hydrogen Peroxide (P)	2	36.4															
37% Formaldehyde (F)	6	-2.1															



**Leverandør**  
 Firma: World Medico ApS  
 Adresse: Østerfælled Torv 3  
 Post nr. / By: 2100 / København Ø  
 Land: Danmark  
 Telefon / E-mail: +45 53692712 / info@worldmedico.dk

